

TEL.+1(562)594-5060 FAX.+1(562)594-5061

TEL.+49(0) 7153 934 291 FAX.+49(0) 7153 934 299

#### Magnescale Co., Ltd.

Headquaters

Shinagawa Intercity Tower A-18F, 2-15-1, Konan, Minato-ku, Tokyo 108-6018, JAPAN

TEL.+81(0)463-92-1011 FAX.+81(0)463-92-1012 : 45 Suzukawa, Isehara-shi, Kanagawa 259-1146, JAPAN Tokyo Office : Shinagawa Intercity Tower A-18F, 2-15-1, Konan, Minato-ku, Tokyo 108-6018, JAPAN TEL.+81(0)3-5460-3574 FAX.+81(0)3-5460-9614 Nagoya Office : 2-35-16, Meieki, Nakamura-ku, Nagoya Aichi, 450-0002, JAPAN TEL.+81(0)52-587-1823 FAX.+81(0)52-587-1848 : 2-14-6, Nishi-Nakajima, Yodogawa-ku, Osaka 532-0011, JAPAN TEL.+81(0)6-6305-3101 FAX.+81(0)6-6304-6586 Osaka Office International Sales Department : 45 Suzukawa, Isehara-shi, Kanagawa 259-1146, JAPAN TEL.+81(0)463-92-7971 FAX.+81(0)463-92-7978

Magnescale Americas Inc. : 5740 Warland Drive, Cypress, CA 90630, USA Magnescale Europe GmbH : Antoniusstrasse 14, 73249 Wernau, Germany

The contents of this literature are as of Jul. 2011 This catalog is printed with soy ink. MGS-LS-1107-FN-C

http://www.mgscale.com



### Laserscale®

Magnescale Co., Ltd.

### Safety

#### No compromise for high-accuracy products



The total quality control system that operates throughout the entire design and production process ensures products with enhanced safety, high quality, and high reliability that match our customers' requirements. The company is certified for length calibration in compliance with the traceability system required by the "Weights and Measures Act," and has been granted ISO 9001 certification, which is the international standard for quality assurance.





Isehara plant is registered to ISO 9001 (Quality

Our products comply with CE Marking requirements, have acquired UL certifications and meet other regulations, ensuring safe use the world over.

We have met:

EMC Directives(CE)

FCC regulation

EMI: EN 55011 Group 1 Class A / 91

FCC Part 15 Subpart B Class A

EMS: EN 61000-6-2

for Products with built-in AC power supply:

•UL 61010-1

for Products with Laser:

•DHHS Class 1 (21CFR1040.10)

### Traceability

Traceability Flow Chart (Length)

National Primary Standards National Institute of Advanced Industrial Science and Technology (AIST)



International Committee for Weights and Measures (CIPM)

International Bureau of
Weights and Measures (BIPM)

Magnescale Corporation

National lodine saturation absorption stabilized



Practical Standards

Secondary

Standards

Stabilized He-Ne Laser (633nm)

He-Ne laser at 633nm





Products

Laserscale Contents Safety Traceability Contents Introduction Principle Application Lineup BS65-R BH25-RE/BH25-NE 16 BH20-RE/BH20-NE 18 BH200-RE/BH200-NE 20 BL57-RE/BL57-NE 22 BL55-RU BD95 Connection Cable 34

\* The product name "Laserscale" is trademark of Magnescale Corporation.

Technology

<sup>\*</sup> When using our devices with machines to which the European Machinery Drirective applies, please make sure that the devices when installed on the machines fulfil the applicable requirements of the Directive.

<sup>\*</sup> Standards or regulations to be complied with may vary by produc



### The world of super-resolution is going further than 1nm

Laserscale easily achieves measurement and control by ultra high resolution less than 1nm.

Sinusoidal wave of approx. 138nm signal period is generated by a hologram scale with high diffraction efficiency and a high resolution detecting head based on grating interference method, strong against disturbance by air pressure or current, plus easy to install (BS series). Signal distortion in principle remains minimal at high S/N ratio. The highest resolution reaches 17pm in combination with a interpolator featuring automatic compensation.

138nm

High-resolution scale with signal wavelength of approx. 138nm outperforms light wave interferometer systems

#### Ultra-high resolution

Volume holography technology of Laserscale® achieves high diffraction efficiency to generate high S/N signal and big output signal.

#### Highest resolution 17pm

One count movement by holographic grating of 550nm wave length diffracts interfering signal

to 4 periods, resulting 1/4 of original signal (approx.  $0.14\mu m$ ). And signals go through electrical Interpolator, it will be maxmum 17pm resolution.

#### Ultra-high resolution and high speed response

Grating interference principle linear encoder generates signal of approx.


0.14 µm period, that is 1/140 of conventional linear encoder with 20 µm signal period.

Also using our interpolator, 17 pm resolution and high response as max.

response speed 400 mm/s are available.

Model	Output	Number of divisions	Resol	Resolutions Maximum response	
BS series	Binary	8000	17	pm	400mm/s
Signal wavelength: 138nm	AB quadrature	32	4.31	nm	60mm/s

#### High stability : Free from humidity, air pressure, or air disturbances



#### Easy installation & maintenance

#### [Easy to handle]

Large tolerance for installation


Easy installation, remarkable for ultra-high resolution, accuracy and non-contact detection

#### No electric adjustment after installation

Even with high tolerance for installation, no electric adjustment required after installation.

#### Protected holographic grating

Holographic grating is protected with cover glasses, which guards the grating against external pressure. The glass can be wiped out to clean dust and dirt.



### Principle

The semiconductor laser beam is split by a polarized light beam splitter into S and P polarized light beams, then diffracted through a volume hologram lattice with very high diffraction efficiency. The two diffracted beams pass through separate 1/4-wavelength plates to a mirror, which reflects the beams back through the plates. This process converts the S polarized beam to P polarized light and the P polarized beam to S polarized light.


The two beams are diffracted again through the volume hologram lattice, then super-positioned

by the polarized light beam splitter to create interference. All interference travels to the photo-detector side due to conversion of the polarization direction. Since double diffraction adds +2 Kx and -2 Kx phases to each beam, the interference is subject to four light-dark inversion cycles for each lattice scale of movement. Thus a lattice pitch of 0.55 µm. produces a signal wavelength of  $0.55/4 = 0.1379 \mu m$ . This detecting optics is free from fluctuations and change in air pressure, since the light path of both left and right changes identiacally even with


the change in wavelength of the optical source.

Repeatability and returning errors do not


occur in principle.




⇒: Direction where light vibrates···Right and left ⊚: Direction where light vibrates···Back and forth



### Application



Ultra high precision air stages (vacuum resistant)




Surface roughness/ contour measuring machines




DUV-based automatic wafer defect classification systems




Non-contact measuring machines



Micro grinders



Aspherical surface machining



# Lineup


		Series	Feature	Max.resolution	Accuracy	Measuring length	Interpolator	Output	Response speed	Page			
				BS78	Low expansion glass	17pm	±0.04μm	10mm~420mm	BD96	40bit Binary	- 400mm/s	P. 10	
RS				C E		BS/8	LOW expansion glass		(Measuring length 40mm)	10(1)[11~420(1)[11	(BD95)	Serial	4001111//5
λ=138nm Transmission					BS65-R	Long-length type	17pm	L<460: (0.1+0.4L/100)μm	160mm~960mm	BD96	40bit Binary	400mm/s	P. 12
		500011	Blue plate glass	17pm	L≧460:3µmp-p L:Measuring length(mm)	10011111~90011111	(BD95)	Serial	40011111/3	1.12			
		DUOS DE AIS	Low expansion glass	0.00405	±0.5μm (30mm-170mm)	Low expansion glass: nm) 30mm~420mm	BD96	40bit Binary	700,000	D.44			
			BH25-RE/NE	Blue plate glass	0.03125nm	±1μm (220mm-420mm)	Blue plate glass: 30mm~420mm	guyag	Serial	- 700mm/s	P, 14		
BH		BH20-RE/NE	RH20-RE/NE	302,400Pulse/rotation 680,400Pulse/rotation	rotation 1.5nrad	_	Radius 12.03mm Radius 27.07mm Radius 36.1mm Radius 41.72mm	BD96	40bit Binary	555min <sup>-1</sup>	P, 16		
$\lambda$ =250nm Reflection			DIESTIENE	907,200Pulse/rotation 1,048,576Pulse/rotation					Serial	(1,190min <sup>-1</sup> , 529min <sup>-1</sup> ) 396min <sup>-1</sup> , 343min <sup>-1</sup> )	1,10		
		BH200-RE/NE	907,200 Pulse/rotation	6.93nrad	r36.1 Scale only	Radius 36.1mm	Nothing	Clock pulse (LVDS)	13,000min <sup>-1</sup>	P, 18			
		BL57-RE	Low expansion glass Blue plate glass	0.1/0.05/0.02/ 0.01μm	±0.5μm (30mm-170mm) ±1μm (220mm-370mm) ±1.5μm (420mm-1,060mm) Please ask for more than 1,060mm	Low expansion glass: 30mm~410mm Blue plate glass:	Built-in I/F Box	AB quadrature	1,500, 650, 300, 120mm/s	P, 20			
	7		, J	0.4μm (1Vp-p) 0.1/0.05/0.02/		60mm~1,060mm	Nothing	Analog	3,000mm/s				
		BL57-NE	Low expansion glass  Blue plate glass	0.1/0.05/0.02/ 0.01μm	±0.5μm (30-170mm) ±1μm (220-370mm)	Low expansion glass: 30mm~420mm Blue plate glass:	Built-in I/F Box	AB quadrature	1,500, 650, 300, 120mm/s	P, 20			
			Dido piate glass	0.4μm (1Vp-p)	±1.5μm (420-1,060mm) Please ask for more than 1,060mm	60mm~1,060mm	Nothing	Analog	3,000mm/s				
$\lambda$ =400nm Transmission	D.	D.	D.	D.	BL55-RU	Blue plate glass -	0.1μm~0.01nm	±2.5μm (70mm-370mm)	Blue plate glass:	Built-in I/F Box	AB quadrature	1,500, 650, 300, 120 mm/s	P, 28
		DLUU-NU	Dine higher Alass -	400nm (1Vp-p)	±4.5μm (370mm or more)	70mm~1,020mm	Nothing	Analog	3,000mm/s	P, 28			
			1			1		1					

**BS78** (with/without reference mark)

High-speed and high-resolution, while maintaining stable, ultraprecision measuring. Ideal for precision stages, semiconductor inspection/manufacturing systems, and ultraprecision processing machines.





- High-resolution scale with signal wavelength of approx. 138nm, outperforming light wave interferometer systems
- High stability, unaffected by humidity, air pressure and air disturbances
- Reference point accuracy: ±0.1μm
- Scale accuracy : ±0.04µm or better (measuring length : 40 mm)
- Non-contact design eliminates return error.
- Special non-magnetic and vacuum-compatible models available
- Using low expansion glass: -0.7 x 10<sup>-6</sup>/°C (measuring length: 10 to 420 mm)



#### **External Dimensions**

#### ● BS78-xxxN(NS)(measuring length:40/120/170/220/370/420)





Note 2:The surface roughness of the scale mounting surface is Rmax = 6.3S.


Note 3:The surface roughness of the detector head mounting surface is Rmax = 12.5S.

Note 4:"M" refers to the machine guide.

Note 5:Reference point detection direction : Standard (Scale movement direction → with the head stationary)

Model	L1	L2	L3	N
BS78-40R (RS)	66	_	_	2
BS78-120R (RS)	146	50	_	6
BS78-170R (RS)	196	75	_	6
BS78-220R (RS)	246	100	_	6
BS78-370R (RS)	396	75	75	10
BS78-420R (RS)	446	100	100	10

#### ● BS78-xxxN(NS)(measuring length:70/270/320)







Model	L1	L2	N	
BS78-70R (RS)	96	_	4	
BS78-270R (RS)	296	120	8	
BS78-320R (RS)	346	120	8	
			Unit: mm	

#### External Dimensions

#### ● BS78-xxx(NS)(measuring length:40/120/170/220/370/420)






Model	L1	L2	L3	N
BS78-40N (NS)	66	_	_	2
BS78-120N (NS)	146	50	_	6
BS78-170N (NS)	196	75	_	6
BS78-220N (NS)	246	100	_	6
BS78-370N (NS)	396	75	75	10
BS78-420N (NS)	446	100	100	10

Unit: mm

#### ● BS78-xxx(NS)(measuring length:70/270/320)

12



Main specifications					
Model	B	S78			
Measuring length	10(only N/NS)/40/70/120/	/170/220/270/320/370/420			
Overall length	58mm (Measuring length 10mm) Measuring length	Single direction			
Max. travel length	Measuring length + 2mm (Measuring length 10mm) Measuring length 10mm) Measuring length 10mm (Measuring length 10mm	58mm (Measuring length 10mm) Measuring length + 26mm (Measuring length 40mm to 420mm)  Measuring length + 2mm (Measuring length 10mm) Measuring length + 10mm (Measuring length 40mm to 420mm)  NS type, RS type:  ± 0.03μm (NS type) ± 0.04μm  ± 0.11μm ± 0.18μm ± 0.25μm  ± 0.08μm ± 0.08μm ± 0.20μm  ± 0.35μm ± 0.50μm ± 0.65μm  ± 0.44μm  ± 0.1μm (Only R/RS type)  and every 50mm from the center to the left and to the right (BS78 models with measuring lengths of 320, 370, 420mm: 20mm offset from the center at 50mm interva			
Accuracy(at 20°C)	± 0.03μm (NS type) ± 0.04μm ± 0.10μm ± 0.18μm ± 0.25μm ± 0.34μm ± 0.39μm	± 0.06μm ± 0.08μm ± 0.20μm			
Reference point accuracy	± 0.1μm (Or	nly R/RS type)			
Reference point position	At the center, and every 50mm from the center to the left and to the right (BS78 models w	ith measuring lengths of 320, 370, 420mm: 20mm offset from the center at 50mm intervals)			
Reference point detection direction	Single	direction			
Return error	This is virtually eliminated. It should be considered to be	less than two resolution limits of the detector that is used.			
Repeatability	This is virtually eliminated. It should be considered to be	less than one resolution limit of the detector that is used.			
Temperature coefficient	-0.7x	110 <sup>-6</sup> /°C			
Light source	Semiconductor laser Wavelength	h, 780nm, Maximum output 10nW			
Radiation power	DHHS	S class 1			
Detection type	Diffraction gra	ating scan type			
Operating temperature	10°C to 30°C (N	No condensation)			
Storage temperature	-10°C to 50°C (Hu	umidity 60% or less)			
Weight	0.4/0.41/0.42/0.43/0.44/0.45/0	0.46/0.47/0.48kg (including head)			

# BS65-R (with reference mark)


High accuracy LASERSCALE™ with built-in optical zero point



- Signal wavelength of 138nm,
- High accuracy, high resolution laser scale:
   Accuracy: L < 460: (0.1+0.4L / 100) μm p-p</li>
   (L=measuring length in mm)
- Resolution : 0.07nm
- High accuracy optical reference point : ± 0.1 μm
- Measuring length: 160 mm to 960 mm / 6.29" to 37.79"
- Easy installation
- Minimal effect from disrupted air current and atmospheric changes.



#### • BS65-xxxN(NS)(measuring length:160/260/360/460/560/660/760/860/960)



Main specifications	
Model	BS65-R
Measuring length	160/260/360/460/560/660/760/860/960
Overall length	Measuring length + 36mm
Max. travel length	Measuring length + 10mm (5mm on each side)
Accuracy (at 20°C)	L < 460 : (0.1 + 0.4L/100) μm p-p L≥ 460 : 3μm p-p L : Measuring length (mm)
Reference point accuracy	±0.1µm
Reference point position	At the center, and every 50mm from the center to the left and to the right (BS78 models with measuring lengths of 320, 370, 420mm: 20mm offset from the center at 50mm intervals)
Reference point detection direction	Single direction
Return error	This is virtually eliminated. It should be considered to be less than two resolution limits of the detector that is used.
Repeatability	This is virtually eliminated. It should be considered to be less than one resolution limit of the detector that is used.
Temperature coefficient	8 x 10 <sup>-6</sup> /°C
Light source	Semiconductor laser Wavelength, 780nm, Maximum output 10nm
Radiation power	DHHS class 1
Detection type	Diffraction grating scan type
Operating temperature	10°C to 30°C (No condensation)
Storage temperature	-10°C to 50°C (Humidity less than 60%)
Weight (kg/lbs)	0.5kg to 1.0kg (including head)


R: with reference point
Effective length



High-accuracy, reflective Laserscale with signal wavelength of 250nm Ideal for low-profile stages, semiconductor back-end processing equipment and precision microscopes




- Signal wavelength: 250nm
- High accuracy : ±1μm/420nm
- High response speed : 700mm/s
- Maximum resolution: 0.03125nm
- Available : with/without reference point
- Completely non-contact design Return error is theoretically eliminated.
- Scale : Blue plate glass/Low expansion glass
- Thin head with thickness of 12mm
- Supporting various resolutions and output modes (Depending on the interpolator connected.)
- Special non-magnetic and vacuum-compatible models available



#### External Dimensions


#### ● BH25-RED BH25-\*\*\*REDB



3. The surface roughness of the detector head mounting surface is Rmax = 6.3S(250winch). 4. "M"refers to the machine guide.

5. Mount and sdjust the paired reference plates so that their reference surfaces have a parallelism of 0.01 or less with respect to the machine guide.

#### ● BH25-RED BH25-\*\*\*REDC



Notes: 1. The items marked by an asterisk indicate the machining dimensions on the mounting surface. 2. The surface rounghness of the scale mounting surface is Rmax = 6.3S (250µinch).

3. The surface rounghness of the detector head mounting surface is Rmax = 6.3S (250µinch). 4. "M\*refers to the machine guide.

5. Mount and sdjust the paired reference plates so that their reference surfaces have a parallelism of 0.01 or less with respect to the machine guide.

		Unit: mr					
Main specifications							
Model	BH25-RED	BH25-NED					
Measuring length (mm)	30/70/120/170/220/270/320/370/420	(Low expansion glass/Blue plate glass)					
Overall scale length	Measuring le	ngth +26mm					
Maximum travel length	Measuring le	Measuring length +10mm					
Accuracy (at 20°C)	±0.5μm (30 to 170mm)	±1.0µm (220 to 420mm)					
Grating pitch	1.0	1.0µm					
Signal pitch	250	inm					
Reference point	With reference point	None					
Direction of starting point detection	For one	None					
Output signal	AB phase and In	tarporator BD96					
Output resolution	AB phase and BD96 connection(De	epend on the number of partitions.)					
Temperature coefficient	-0.7 x 10 <sup>-6</sup> /°C (Low expansion gla	ass) 8 x 10 <sup>-6</sup> /°C (Blue plate glass)					
Light source	Semiconductor laser Wavelength	, 780nm, Maximum output 6mW					
Detection system	Diffraction grating	scanning system					
Operating temperature range	+10°C to +30°C (	No condensation)					
Storage temperature range	-10°C to +50°C (Hum	nidity less than 60%)					
Maximum response speed	700mm/s (When co	nnected with BD96)					




Compact, reflective rotary Laserscale featuring high accuracy, high resolution and high response speed.

Ideal for high-resolution angle measuring in HDD manufacturing equipment, precision measuring instruments, and aspheric surface processing machines.



#### **External Dimensions**

#### ● BH20-NED




Main specifications					
Detection head					
Model	BH20-RED	BH20-NED			
Detection system	Diffraction grating	g scanning system			
Light source	Semiconductor laser with power of 6	mW or less and wavelength of 790nm			
Reference point	Single reference point; single -direction detection	None			
Grating pitch	1.0µm				
Signal pitch	250	250μm			
Reference point	With reference point	None			
Direction of starting point detection	For one	None			
Scale		nm (scale : Ni-Co or Si ) 32, 27.073, 36.097mm (rotary scale Ni-Co)			
Maximum response speed	1,500mm/s(When using analog output),	700mm/s(When using interpolator BD96)			
Operating temperature range	+10°C to +30°C no condensation. A	Avoid operating under high humidity.			
Storage temperature range	.0°C to .50°C no condensation A	void aparating under high humidity			

#### Film scale (BE10)

Detection radius		12.032mm 27.073mm		36.097mm	41.723mm
Detection track	Max.radius	13.532mm	28.573mm	37.597mm	43.523mm
Detection track	Min.radius	10.532mm	25.573mm	34.597mm	40.597mm
External form	Internal diameter	7.7mm	19.77mm	29.00mm	34mm
External form	External diameter	13.45mm	28.57mm	37.60mm	43.523mm
Scale diameter		20ømm	59ømm	77ømm	87ømm
Number of output of one rotation	302400		680400	907200	1048576
Maximum respon	nse speed *	1190 min <sup>-1</sup>	529 min <sup>-1</sup>	396 min <sup>-1</sup>	343min <sup>-1</sup>

 $<sup>^{\</sup>star}$  When using cable length 1m and Analog output. However, the Max.response speed is limited depending on the cable length.




# BH200-RE / BH200-NE (with reference mark) / (without reference mark)


Compact, reflective rotary Laserscale featuring high accuracy, high resolution, and high-speed response, Ideal for high-resolution angle measuring in HDD manufacturing equipment and precision measuring instruments

(signal wavelength: 250nm)



#### External Dimensions





±70μm(output ±40%)

±50μm(output ±40%)

CLK signal(LVDS), 1/2 or 1/4 CLK signal(LVDS)<sup>2</sup>

Switch over 1/2 or 1/4(TTL)

DC ±5V(±5%)

DC +5V: 400mA, DC -5V: 200mA +10 to +30 °C (head), 0 to +50 °C (detector)

-10 to +50 °C No condensation Avoid operating under high humidity

0.5 nsec (@5000rpm)

BH200-NE

\*1 Zero point signal is asynchronous to CLK, 1/2CLK and 1/4CLK signals. Detection is unidirectional.

\*2 1/2 or 1/4 signals to CLK signal frequency. No output with input frequency of 50MHz or less. 1/2 CLK and 1/4 CLK signals cannot be used simultaneously with CLK signal. \*3 Jitter of CLK signal: Pulse duration variation at 1000 pulses(3 or p-p). Measured by inspection equipment at Magnescale Corporation.

Head position

Output signal

Input signal

Power supply

Jitter (target)<sup>-3</sup>

Maximum power consumption

Optical fiber minimum bend radius

Operating temperature Storage temperature

ΔΥ

ΔZ

# BL57-RE / BL57-NE (with reference mark) / (without reference mark)

Supports a wide range of applications and offers the highest performance in its class. Ideal for precision stages, semiconductor inspection systems, precision processing machines, and liquid crystal manufacturing equipment.



#### BL57-RE

- Achieves a measuring length of up to 1,360mm upon request, and offers the highest-level response speed and accuracy in its class.
- With a head signal pitch of 400nm, interpolation error is virtually non-existent.
- Built-in Reference point.

Applications: Precision measuring equipment, precision stages.


#### BL57-NE


- Compact size makes machine integration much easier
- Theoretically unaffected by changes in temperature, humidity, air pressure and air movement. Unparalled measuring stability achieved by use of low expansion glass
- With a head signal pitch of 400nm, interpolation error is virtually non-existent. Applications: High-accuracy microscopes, stages, measurement equipment.

Type example : BL57-160REFB B: Blue plate glass: C: Low-expansion glass A: 4-split A/B phase output A: 4-split 8-split A/B phase output G: 20-split 40-split A/B phase output H: Analog 1Vp-p output E: Open type scale R: with Reference point; N: without Reference point Effective length

#### External Dimensions


#### • BL57-xxxRE\*B (Effective length: 006/016/026/036/046)






+0.05 Reference hole	.0.05
98	ø4 <sup>+0.05</sup> Reference hole
88 5-05	50 3-0
0 Defended a late W (4.5)	21 42 00 00 00 00 00 00 00 00 00 00 00 00 00


Reference plate (t=5) For BL57-006RE\*B, BL57-026RE\* For BI 57-016RF\* B, or BL57-046RE\*B B or BL57-036RE\*B







For BL57-066RE\*B, BL57-096RE\* B, or BL57-106RE\*B

#### • BL57-xxxRE\*B (Effective length: 056/066/076/086/096/106)








Model	Effective length	L1	L2	L3	L4	NA	SC	SP	Р
BL57-056RE*B	560	596	100	175	75	2	8	6	16
BL57-066RE*B	660	696	75	225	75	3	9	7	18
BL57-076RE*B	760	796	100	250	75	3	10	8	20
BL57-086RE*B	860	896	100	250	75	4	12	10	24
BL57-096RE*B	960	996	75	300	75	5	13	11	26
BL57-106RE*B	1060	1096	75	300	75	6	15	13	30

- Note 1: The items marked by an asterisk indicate the machining dimensions on the mounting surface. Note 2: The surface roughness of the scale mounting surface is Rmax = 6.3 S (250 µinch). Note 3: The surface roughness of the detector head mounting surface is Rmax = 12.5 S (500µinch). Note 4: "M" refers to the machine guide.
- Note 5: When mounting the reference plate (reference plate W), adjust the plate so that the parallelism between the corresponding scale abutting surface and the machine guide is 0.01mm or less.

Reference plate W (t=2.2)

For BL57-006RE\*C, BL57-026RE\*C, or BL57-031RE\*C

#### • BL57-xxxRE\*C (Effective length: 003/006/011/016/021/026/031/036/041)



360 396 75 75 5 4 10 410 446 100 100

296 120

146

196

346 120

Note 1: The items marked by an asterisk indicate the machining dimensions on the mounting surface. Note 2: The surface roughness of the scale mounting surface is  $R = 6.3 \, \text{S} \, (250 \, \mu \text{inch})$ .

Stopper (ABS)

Spacer (t=2.2)

Note 3: The surface roughness of the detector head mounting surface is Rmax = 12.5 S (500µinch). Note 4: "M" refers to the machine guide.

Reference plate W (t=2.2)

For BL57-003RE\*C, BL57-011RE\*C, BL57-016RE\*C, BL57-021RE\*C, BL57-036RE\*C, or BL57-041RE\*C

Note 5: When mounting the reference plate (reference plate W), adjust the plate so that the parallelism between the corresponding scale abutting surface and the machine guide is 0.01mm or less.

Main specifications [BL57-RE]						
Model		F	G	Н		
Output sign	al form	A/B quadra	ature output	Analogue output		
Detection sy	ystem	Diffra	ction grating scanning s	ystem		
Scale length	Measuring length(mm)	30. 60. 1	10. 160. 210. 260. 310.	360. 410		
(Low expansion	Maximum movable length	Measuring I	length + 10mm (5mm o	n each side)		
glass)	Entire scale length	N	Measuring length + 36m	m		
Scale length	Measuring length(mm)	60. 160. 260. 3	60. 460. 560. 660. 760.	860. 960. 1060		
(Blue plate	Maximum movable length	Measuring	Measuring length +10mm (5mm on each side)			
glass)	Entire scale length	N	Measuring length + 36m	m		
Grating pitch		1.6µm				
Signal pitch		0.4µm				
Output signal		Differential (compliant with EIA-422)		Differential (only zero point output models are compliant with EIA-422)		
Resolution		0.1/0.05μm (switchable)	0.02/0.01µm (switchable)	0.4μm (1Vp-p)		
Accuracy (a	t 20°C)	±0.5μm(30 to 170mm) / ±1.0μm(220 to 370mm) / ±1.5μm(420mm or more)				
Thermal exp	oansion coefficient	Low expansion gla	ass:-0.7x10 <sup>-6</sup> /°C •Blue p	late glass:8x10°/C		
		1,500mm/s(0.1μm) 650mm/s(0.05μm)	300mm/s(0.02μm) 120mm/s(0.01μm)	3000mm/s		
Maximum re	esponse speed	Minimum phase difference:38ns	Minimum phase difference:38ns	Max 7.5MHz		
			<b>***</b>			

Model		F	G	Н	
Alarm		High impedance, alarm by output signal when maximum response speed is exceeded or signal level error detected		None	
Reference p	oint position	User definabl	le (within the range of eff	fective length)	
Reference poi	int accuracy (at 20°C)	±0.4μm (deper	nding on machine mover	ment accuracy)	
Reference point output signal		Unidirectional synchronous reference point (specify the position and detection direction)			
	Cable length	1m (Note 4)			
Head cable	Bending radius	When stationary : 10mm			
Output cable	e length	15m Max (Note 2)(to the	15m Max(Note1) (Note 2		
Power suppl	ly (Note 3)	+5V (±5%)			
Power consu	umption	450mA (no load) 600mA (with 120 ohm termination)			
Vibration res	istance	100m/s² (50 to 2000Hz)			
Impact resis	tance	200m/s²			
Operating te	mperature range	0 to +40°C(No condensation)			
Storage tem	perature range	-10 to + 50°C			
Light source		Semiconductor laser with power of 4mW and wavelength of 790nm			
Radiation po	wer	JIS Class 1 equivalent, DHHS Class 1 equnivalent			
Cable length	(m)	Maximum response speed (mm/s)			
3		3000			
0		0000			

1660

BL 57-011RF\*C

BI 57-016RF\*C

BL57-021RE\*C

BL57-026RE\*C

BL57-031RE\*C

BL57-036RE\*C

BL57-041RE\*C

110

160

210

260

310

Note 1: There is a correlation between the maximum response speed and output cable length (the part beyond the interface box). Note 2: A power supply line longer than 10m is incompatible with EN61000-6-2. Take surge protection measures upon use.

Note 3: Satisfy the required specifications at the connector input section.


Note 4: Special models can support up to 3m. However, the maximum response speed is limited depending on the cable length.(In a 3m cable, the maximum response speed is two-thirds that of a 1m cable.)
Note 5: Special models can support up to 3m. However, the maximum response speed is finited depending on the cable length.(In a 3m cable, the maximum response speed is two-thirds that of a 1m cable.)

#### **External Dimensions**


Reference plate W (t=5

For BL57-006NE\*B, BL57-026NE\*B, or BI 57-046NF\*B

#### ● BL57-xxxNE\*B (Effective length: 006/016/026/036/046)




Stopper (ABS)



Unit: mm

#### ● BL57-xxxNE\*B (Effective length: 056/066/076/086/096/106)

For BL57-016NE\*B or BL57-036NE\*B



BL57-056NE\*B

BL57-066NE\*B

BL57-076NE\*B

BL57-086NE\*B

BL57-096NE\*B

BL57-106NE\*B

560

860

960

596 100 175 75 2

996 75 300 75 5 13

1060 | 1096 | 75 | 300 | 75 | 6 | 15 | 13 | 30

896 100 250 75 4 12 10 24

796 100 250 75

11 26

Note 1: The items marked by an asterisk indicate the machining dimensions on the mounting surface. Note 2: The surface roughness of the scale mounting surface is  $Rmax = 6.3 \text{ S} (250 \mu \text{ inch})$ .


Reference plate W (t=5)

Note 3: The surface roughness of the detector head mounting surface is Rmax = 12.5 S (500µ inch). Note 4: "M" refers to the machine guide.

Note 5: When mounting the reference plate (reference plate W), adjust the plate so that the parallelism between the corresponding scale abutting surface and the machine guide is 0.01mm or less.

Stopper (ABS)

#### • BL57-xxxNE\*C (Effective length: 003/007/012/017/022/027/032/037/042)



Reference hole 98 88 5-05	Reference hole 50 3-05	2.8 (2.8) (2.8) (3.8) (4.1) (4.1) (5.1) (6.1) (7.1	2-05
<u>42</u>	60   0	Ø2.8 <sup>+0.05</sup> hole	Spacer (t=2,2)
Reference plate W (t=2.2)	Reference plate W (t=2.2)	Stopper (ABS)	opacoi (t-LiL)
For BL57-007NE*C, BL57-027NE*C, or BL57-032NE*C	For BL57-003NE*C, BL57-012NE*C, BL57-017I BL57-022NE*C, BL57-037NE*C, or BL57-042N	NE*C, E*C	

Model	Effective length	L1	L2	L3	SC	SP	Р
BL57-003NE*C	30	56	_	_	1	_	2
BL57-007NE*C	70	96	_	_	2	_	4
BL57-012NE*C	120	146	50	_	3	2	6
BL57-017NE*C	170	196	75	_	3	2	6
BL57-022NE*C	220	246	100	_	3	2	6
BL57-027NE*C	270	296	120	_	4	2	8
BL57-032NE*C	320	346	120	_	4	2	8
BL57-037NE*C	370	396	75	75	5	4	10
BL57-042NE*C	420	446	100	100	5	4	10

Note 1: The items marked by an asterisk indicate the machining dimensions on the mounting surface. Note 2: The surface roughness of the scale mounting surface is Rmax = 6.3 S (250µ inch).

Note 3: The surface roughness of the detector head mounting surface is Rmax = 12.5 S (500µ inch). Note 4: "M" refers to the machine guide.

Note 5: When mounting the reference plate (reference plate (W), adjust the plate so that the parallelism between the corresponding scale abutting surface and the machine guide is 0.01mm or less.

Main specifications[BL57-NE]								
Model		А	F	G	Н			
Output sign	al form	A	B quadrature outp	out	Analogue output			
Detection sy	ystem		Diffraction grating	scanning system				
Scale length	Measuring length(mm)	30.	30. 70. 120. 170. 220. 270. 320. 370. 420					
(Low expansion	Maximum movable length	A F G  Orm  A/B quadrature output  Diffraction grating scanning system  30. 70. 120. 170. 220. 270. 320. 370. 4  wimum movable length  Measuring length +10mm (5mm on each strescale length  Measuring length +26mm  60. 160. 260. 360. 460. 560. 660. 760. 860. 9  wimum movable length  Measuring length +10mm (5mm on each strescale length)  Measuring length +10mm (5mm on each strescale length)  Measuring length +10mm (5mm on each strescale length)  1.6μm  0.4μm  Differential (compliant with EIA-422)  0.1μm  0.1/0.05μm (can be changed)  ±0.5μm (30 to 170mm)/ ±1.0μm (220 to 37 ±1.5μm (420mm or more)  Low expansion glass: -0.7 x 10-6/°C · Blue plate glassing stress of the s	side)					
glass)	Entire scale length		Measuring le	ngth + 26mm				
Scale length	Measuring length(mm)	60. 160. 2	260. 360. 460. 560	0. 660. 760. 860.	960. 1060			
(Blue plate	Maximum movable length	Meas	suring length +10n	nm (5mm on each	side)			
glass)	Entire scale length		Measuring length + 36mm					
Grating pitc	h	1.6µm						
Signal pitch		0.4μm						
Output sign	al	Differenti	al (compliant with	EIA-422)	Differential			
Resolution		0.1µm			0.4μm (1Vp-p)			
Accuracy (a	it 20°C)	±0.5μn			70mm)/			
Temperature	expansion coefficient	Low expansion	on glass: -0.7 x 10	6/°C+Blue plate gla	ass:8 x 10 <sup>-6</sup> /°C			
		1000mm/s			3000mm/s (Note 1)			
Maximum r	response speed				Max 7.5MHz			
				ı	<b>***</b>			


Model		A	F	G	Н	
Alarm		High-impedance A/B quadrature output signals when signal level error detected.	High-impedance output when maximum response speed exceeded or signal level error detected.		None	
Head	Cable length		300	min		
cable	Bending radius		When statio	nary: 10mm		
Output cable length		15mMax (Note 2) (to the electronic control section)			15mMax (Note 1) (Note 2)	
Power source (Note 3)		+5V (+10%-5%)				
Power consumption		200 mA (no load) 250 mA (9120 ohm termination)	290mA 350mA (120 c	250 mA (no load,120 ohm termination)		
Vibration resis	stance	100m/s²(50 to 2000Hz)				
Impact resista	ance	200m/s²				
Operating ten	nperature range	0 to +40°C(no condensation)				
Storage temp	erature range	-10 to + 50°C				
Light source		Semiconductor laser with power of 4mW and wavelength of 790nm				
Radiation pov	ver	JIS Class 1 equivalent, DHHS Class 1 equivalent				
Cable length (m)		Maximum response speed (mm/s)				
3		3000				
9		2330				

1660

Note 1: There is a correlation between the maximum response speed and output cable length (the part beyond the interface box).

Note 2: A power supply line longer than 10m is incompatible with EN61000-6-2. Take surge protection measures upon use. Note 3: Satisfy the required specifications at the connector input section.

### **BL57-RE** supporting vacuum environment (Special models)




#### Vacuum-compatible, open type with reference point. Allowing ultra-precise positioning in a vacuum environment.

- Ultimate vacuum of 10⁻⁵ Pa class.
- Emitted gas flow rate of 10<sup>-6</sup> Pa·m³ class.
- Signal pitch 0.4μm
- Built-in reference point.

Applications: Semiconductor inspection systems, length measuring SEM.

\*1: For dimensions of head, scale, and I/F box, see the page on BL57-RE. \*2: Cable length in vacuum and in atmosphere (L1 + L2) is up to 3m.



### BL55-RU (with reference mark)

Sealed type linear encoder with non-contact design offers minimal reversal error. The enclosed design protection against contaminates (IP53)




- All the advanced features of Magnescale's Laserscale technology have been incorporated into the environmentally protective design.
- Highest response speed and accuracy in its class.
- Non-contact design of the detector head eliminates inherent mechanical error, and achieves a repeat accuracy of 0.1 µm or less.
- Given the head signal pitch of 400nm, the interpolation error is practically non-existent.
- Built-in reference point.

Applications: Precision measuring equipment precision stages.




#### External Dimensions



Note 1: Parallelism is 0.05mm or less when Measuring length is 120mm or less. Note 2: Parallelism is 0.1mm or less when Measuring length is 170mm or less. Note 3: M indicates the machine guide (machine movement).

Main sp	ecifications					
Model		F	G	Н		
Output sign	nal form	A/B quadra	ature output	Analogue output		
Detection s	ystem		Diffraction grating scanning system			
Scale length	Measuring length (mm)	70 • 120 • 170 • 2	220 • 270 • 320 • 370 • 420 • 470 • 520 • 570 • 620 • 720 • 77	0.820.870.920.1020		
(Blue plate glass)	Maximum movable length		Measuring length + 2mm			
5)	Overall scale length		Measuring length + 137mm			
Accuracy (2	20°C)		±2.5 (70 to 320mm ) ±4.5μm (370mm or more	)		
Linearity (N	ote 2)		±2.5µm (370mm or more)			
Grating pito	ch		1.6µm			
Signal pitch	ı		400nm			
Output sign	nal	Differential (comp	liant with EIA-422)	Differential (only reference point output models are compliant with EIA-423		
Resolution		0.1/0.05µm(switchable using a switch) (Note 1) 0.02/0.01µm(switchable using a switch)		0.4μm(1Vp-p)		
Repeatabili	ty		0.1μm or less			
Return erro	r		0.1μm or less			
Reference po	oint accuracy (at 20°C)		±0.4µm (depending on machine movement accur-	n (depending on machine movement accuracy)		
Reference p	point position		User definable			
Direction of re	eference point detection		For one			
Temperature	expansion coefficient		8x10 <sup>-6</sup> /°C			
Light sourc	е	Two s	semiconductor lasers with power of 6mW and wavelen	gth of 790nm		
Radiation p	ower		JIS Class 1 equivalent, DHHS Class 1 equivaler	nt		
Operating t	emperature range		0 to +40°C (no condensation)			
Storage ten	mperature range		-10 to + 50°C			
Maximum r	esponse speed	F: 1,500mm/s (0.1 \( \mu m \)) 650mm/s (0.05 \( \mu m \)) Minin G: 300mm/s (0.02 \( \mu m \)) 120mm/s (0.01 \( \mu m \)) Minin		3000mm/s (Note 3) Max 7.5MHz		
Alarm			output signal when maximum od or signal level error detected	None		
	Cable length		1000m (Note 4)			
Head cable	Bending radius		When stationary : 30mm When in motion : 100m	ım		
Output cab	le length	15m	Max	15m Max (Note 3)		
Power sour	ce		+5V (±5%)	·		
Power supp	oly	450mA (no load) 600mA (maximum when cable is connected)				
Protective of	design		IP53 or equivalent (when air is supplied : IP64 or equ	ivalent)		
Vibration re	sistance		100m/s² (50 to 2000Hz)			
Impact resis	stance		200m/s <sup>2</sup>			



- Note 1: Special modes can support AB quadrature output with  $0.01 \mu m$  resolution.
- Note 2: The linearity is the range of scattering when scale accuracy slope is set to zero .

  Note 3: Please inquire for details regarding the correlation between the maximum response speed and the output cable length.
- Note 4: Special models can support up to 3m .

Maximum resolution of 17pm when combined with the BS series. Supporting various serial and binary outputs.




- Maximum resolution : 0.4nm (When connected with BL series) 31pm (When connected with BH series) 17pm (When connected with BS series)
- High response speed: 1,100mm/s (When connected with BL series) 700mm/s (When connected with BH series) 400mm/s (When connected with BS series)
- Various serial or binary outputs

- Includes automatic signal compensation
- AB quadrature output (standard: 4 divisions) (binary output axis 1 or 2 type) BS series: 34.5nm, BH series: 62.5nm, BL series: 100nm
- Maximum no. of divisions: 8000 (When connected with BS and BH series)

#### External Dimensions

### ●BD96-B1 ●BD96-B2 6-M3 Tap (for mounting) 6-M3 Tap (for mounting) (both sides) Base (opposite side) 4-M3 Tap (for mounting) /- Base (opposite side) 4-M3 Tap (for mounting) ●BD96-Y2/M2 ●BD96-Y1/M1 6-M3 Tap (for mounting) (both sides) 6-M3 Tap (for mounting) (both sides) Base (opposite side) 4-M3 Tap (for mounting) Base (opposite side) 4-M3 Tap (for mounting)



\*2 External dimensions of the 1- and 2-axis are identical.

Main specifications	
Model	BD 96
MAX. resolution	0.4nm (When connected with BL series), 0.03125nm (When connected with BH series), 0.017nm (When connected with BS series)
Response speed	400mm/s (When connected with BS series),700mm/s (When connected with BH series),1,100mm/s (When connected with BL series)
MAX. division	025 : 256, 051 : 512, 040 : 400, 050 : 500, 100 : 1000, 200 : 2000, 400 : 4000, 800 : 8000
Alarm	When exceeding the maximum response speed or when the laser signal level is too low (disconnection); LED lights up
Input signal compensation	DC offset, amplitude, phase
Power supply	DC +5V±5% DC +12V±5% DC -12V±5%
Current Consumption ( When scale is connected )	DC +5V: 0.4A DC +12V: 0.7A DC -12V: 0.5A ( 2 axes type )
Operating temperature range	0 to +40°C
Storage temperature range	-10 to +50°C
Dimensions	236 (M) x 215.2 (D) x 48 (H)mm
Weight	Approx. 1.6kg

Shape C: Case type

Scale type S: BS series H: BH series L: BL series

Division 025: 256 divisions 051: 512 divisions 040: 400 divisions 050: 500 divisions 100: 1000 divisions 200: 2000 divisions 400: 4000 divisions 800: 8000 divisions

Axis type 1: 1 axis 2: 2 axes J: Supports long scales of the BLseries (Max. 512 divisions)

Output mode B: Binary (Axis type 1 or J: 40 bits, 2: 20bits) Y: Yaskawa Electric serial M: Mitsubishi Electric serial F: FANUC\*2 serial

<sup>\*</sup> Please inquire about various specifications, such as the number of divisions.

# BD95 Interpolator for Laserscale ™


Interpolator with A/B quadrature output that achieves resolution to 4.3nm.



- High resolution: 4.3 to 34.5nm (depends on the number of splits)
- High response speed: 400mm/s
- DC offset, gain, phase automatic conditioning
- 32 bit binary output by data request input (T14, T16, T17)

#### External Dimensions

#### ● BD95-T10,T13,T14,T15,T16,T17commonness



Linit: mr

Main specifications							
Model	BD95-T13	BD95-T14	BD95-T15	BD95-T16	BD95-T10	BD95-T17	
Resolution (selectable)	34.5 nm (4divisions) or 17.2 50 nm during pito			m (8divisions) or 8.6 nm(16 divisions) 100 nm, 50  8.6 nm (16divisions) or 4.3 nm(32 divisions) 1  50nm, or 10 nm during pitch compensation			
Max. response speed	400 mm/s (with 4 divisions)	275 mm/s (with 8 divisions)	275 mm/s (with 8 divisions)	120 mm/s (with 16 divisions)	120 mm/s (with 16 divisions	) 60 mm/s (with 32 divisions	
Output signal		AB quadrature 1 with / without pitch compensation (compliant with EIA-422) AB quadrature 2 without pitch compensation (compliant with EIA-422) Alarm (compliant with EIA-422) (Switching between automatic reset and holding is possible) LASERSCALE signal (SIN/COS) 32-bit binary data (-T14, -T16 only)					
Alarm		Max. response speed exceeded Low laser signal level(cable broken or disconnected) LEDs (Turn on independently for speed alarm and level alarm) Output signal: Output when either a speed or level alarm occurs. Switching between automatic reset and holding is possible					
Pitch compensation function			AB quadrature 1 only A round-	off error of 1 resolution occurs	S.		
Input signal compensation (On/Off switching is possible)		DC offset, Amplitude leve	el, Phase. Frequencies allowing	g compensation update: Input	signals of 180 kHz or less		
Power supply			DC + 2	4V ± 1V			
Consumption current (when scale is connected)			400mA (r	maximum)			
Operating temperature			0°C to 50°C /	32° F to 122°F			
Storage temperature			-10°C to 60°C	/ 14° F to 140°F			
Dimensions		172 (W)x144(D)x32(H) mm/6.77"(W)x5.66"(D)x1.25"(H)					
Weight			Approx. 0.8 kg	/Approx.1.76lbs			

### Connection Cable

Sc	ales	Extension Ca	ablo*9	Interpolator
Model	Head cable length*1	EXTENSION OF	interpolator	
BS78 BS65-R	2m (Standard)	Robot cable:CK-T61 (1.0m) CK-T24 (3.0m) CK-T54 (6.0m)		BD95
	3m (Standard)	Robot cable:CK-T133 (0.1m) CK-T137 (3.0m) CK-T112 (5.0m)		
BH20-NE	1m (Standard)	Robot cable:CK-T148 (3.0m)		BD96
BH25-RED BH20-RED BL57-RED	1m (Standard)	Robot cable:CE20-03T10 (3.0m) CE20-06T01 (6.0m) CK-T144 (9.0m)		

<sup>\*1</sup> Please ask for other length. \*2 can lengthen to max. 9m. Please ask for more than 9m.

Scales		Futuration C	Interpolator	
Model	Head cable length*1	Extension C	interpolator	
BL55-RU	1m (Standard)	Robot cable:CE20-03 (3.0m) CE20-05 (5.0m) CE20-07 (7.0m)		
BL57-NE (A/B phase)	0.3m (Standard)	Robot cable:CE20-03T07 (3.0m)		Built-in I/F
BL57-RE (A/B phase)	1m (Standard)	CE20-05T05 (5.0m)		Box
BL57-NE (Analog)	0.3m (Standard)	Robot cable:CE20-07T03 (7.0m)		
BL57-RE (Analog)	1m (Standard)	nobot cable.oL20-07 103 (7.011)		

<sup>\*1</sup> Please ask for other length.


The robot cable minimum bend radius: R80mm is fixed repeatedly R10mm.

## Technology

#### Before use

#### Mounting Direction BL55-RU

Please see the diagram blow about an instrallation method of Laserscale.



#### Checking the Installation Direction


Check that the scale is installed in the positional relationship shown in Fig. 3-1.

Except when installed on a vertical axis, only the orientation in Fig. 3-1 should be used.


#### [Note]

If installing on a machine tool or other equipment where powder and dust occur, install using the A orientation since this allows usage of cutting fluid and prevents the intrusion of cutting dust.

Install using the B orientation only in other situations when virtually no foreign objects can enter the scale.







① Slider

The slider has a built-in detector head. The slider is secured in place by the slider holders at shipping.

#### ②Scale uni

The scale unit incorporates a high-accuracy LASERSCALE. It is protected by an aluminum cover.

#### 3 Foot plate

This is used to secure the scale in place.

The number of attached foot plates varies depending on the scale measuring length.

Measuring length of 470 mm or less ..... none

Measuring length of 520 mm to 920 mm ......1

Measuring length of 1020 mm .....

#### 4 Air inlet

This is used when air is injected. To inject air, remove the hex.socket-head set screws covering the inlets, and then attach the hex. sockethead half-union.

\*Please refer to page 39 for an electric supply.


#### Range of Movement Settings BL55-RV

As shown in Fig. 3-3, when the head cable is drawn out to the right side,

the point where the distance from the slider left end and scale unit left end is 28.5 mm is considered the left end movable limit.

Although the slider or scale unit can move an amount equal to the measuring length from this left end movable limit, there is almost no margin at both ends. Be particularly careful that the machine movable length falls within the scale measuring length.

\* The detector head will be damaged if the slider or scale unit is moved past the measuring length. A mechanical limiting mechanism (such as a stopper) is needed for machines that exceed the measuringlength (movable range) of the scale.
Be sure to install this type of device before installing the scale.



#### Compensation Value BS78, BS65

The hologram grating pitch of the scale is approximately 0.55  $\mu$ m/22  $\mu$  inch, while the detector signal pitch is theoretically 1/4 of this, that is 138nm/5.5  $\mu$  inch. This value differs slightly from scale to scale. Therefore, it is necessary to additionally make the pitch compensation for each scale, in order to detect precise displacements.

The lowest four digits of the detector signal pitch for each scale PS = 0.1379  $\square$   $\square$   $\square$   $\mu$ m at 20 °C/68 °F show the compensation value.

Before operation, be sure to always set the compensation value of the scale for the Intarporator when using a Intarporator that allows input of compensation values.

Linear compensation must be performed using the customer's machine if the compensation value cannot be set.

#### When you operate the unit in an environment where the temperature is not 20 °C/68 °F

To operate the unit in an environment where the temperature is not 20 °C/68 °F, adjust the compensation value as shown below to make a temperature adjustment. The following example shows the adjustment method for an operating temperature of 23 °C/73 °F.

Example: The compensation value shown on the compensation value label is 1234.

 $0.13791234 \times \{1 + (23-20) \times (-0.7 \times 10^{-6})\} = 0.13791205$ 

Temperature difference from 20 °C/68 °F -

Coefficient of thermal expansion of the scale —

The compensation value after temperature adjustment will be 1205. Set the detector to this value.


#### Air supply

#### Air Injection Procedure BL55-RU

Chips, cutting oil, and other substances generated by cutting can frequently be scattered in the area around the scale unit, especially when it is installed on a machine tool. Even when not using a machine tool, air should be injected to the scale when installed on machines generating dust or when dust is prevalent in the operating environment.

There are a total of three air inlets, one each on both sides of the scale unit and on the slider.

Supply air to all of the air inlets.




#### Air Pressure Source • Supply Amount BL55-RU

pressure source, the air should always pass through a filter (5  $\mu m)$ , mist separator (0.3  $\mu m)$ , and regulator to remove any dust, mist, and other foreign substances. Sufficient care should be taken since foreign substances

Whenever air is supplied to the scale from an air

entering inside the scale can cause a breakdown of the scale. Air should be supplied to the scale at a pressure of 19.6 kPa per scale. Note that the pressure of the supply section may be reduced due to the pipe length or layout even if the setting at the regulator is 19.6 kPa.




#### Reference point detection direction

The optical built-in reference point (reference mark) of the laserscale can be detected by single direction. Forward detection is set as standard, but it can detect signal from reverse direction depending on the equipment in use. The direction should be specified before order.

Please contact us for further information.

\* Do not detect the reference point from the wrong direction inorder to keep the reliability of the reference point and to avoid deterioration




#### Scale Signal Output

#### AB Signal and Alarm Output Specifications (For output formats F and G) BL55, BL57

- The output specifications are compliant with EIA-422. A signal
- AB signal minimum phase difference t: 38 ns

#### Mote

- An error of about 38 ns is generated due to the synchronization of the AB signal by the 26.3 MHz internal clock.
- The minimum phase error can vary depending on the length of the output cable, cable capacity, receiver load, and other factors.



#### **Connection Specifications**

#### AB Signal Output Type

The line driver used by Magnescale Corporation is compliant with EIA-422.

Also, based on the EIA-422 standards, the common mode voltage between the line driver and line receiver is stipulated as  $\pm$  12 V.


(Using the scale when the common mode voltage of ± 12 V is exceeded can damage the scale.)

To prevent problems between the control devices connected to this Magnescale product, it is recommended that you connect (shared connection) the signal ground (power supply ground) and set the load resistance to 120  $\Omega$ .


Twisted pair cables (1 turn/1 inch min.) with a core thickness of at least AWG28 are recommended for the differential signal cables.

(It is even better if the characteristic differential impedance is the same as the load resistance value.)









#### Analog Output Specifications (For output format H) BL55, BL57

#### SIN/COS output specifications

(Over the entire length and the entire operating temperature range)

Item	Symbol	Spe	ecificati	ons	Units	Remarks
item	Symbol	Min.	Тур.	Max.	UTILS	nemarks
Output signal amplitude	(+VA) - (-VA), (+VB) - (-VB)	0.6	1	1.2	Vp-p	Note 1
Output signal phase difference		80	90	100	deg	
Center voltage	+VOA, +VOB, -VOA, -VOB	2.3	2.5	2.7	V	
Offset voltage	(+VOA) - (-VOA), (+VOB) - (-VOB)	-50	0	50	mV	
Gain unbalance		-6	0	6	%	System 1
Load resistance			120		Ω	

Note 1: When terminator Z0 = 120 $\Omega$  supply voltage= 5V  $\pm$  5% (voltage of load resistance at both ends)

System 1: A signal output voltage p-p value - AB signal output average

AB signal output average x100


where


AB signal output voltage p-p value B signal output voltage p-p value B signal output voltage p-p value 2

#### Output waveform diagram

(when each output is viewed based on 0 V)

The A signal corresponds to SIN, and the B signal corresponds to COS.






#### Reference point output specifications

The output specifications are compliant with EIA-422.

(Over the entire length and the entire operating temperature range)


14	O made at	Specifications			Units
Item	Symbol	Min.	Тур.	Max.	Office
"H" level output	Vozh	2.5	3.4	5	V
"L" level output	VozL	0	0.3	0.5	V



### Reference point signal and SIN and COS signal phases


Item	Specifications		
	Min.	Тур.	Max.
Reference point signal width (Lz)	0.32 μm	0.4 μm	0.48 μm
Position of reference point signal edge a with respect to SIN signal	0°		90°

#### Reference point waveform diagram



#### Analog Output Type

Example of input circuit



 $U0=U1=2.5V \pm 0.2V \quad Z0=120\Omega$ 

Recommended elements

SIN and COS: Differential receiver LMH6654

 $R1 = R2 = 10 \text{ k}\Omega$ 

REF: DS34C86

#### Input/Output Connectors

#### Connectors BL55

#### Interface unit side:

R04-R12M (manufactured by TAJIMI ELECTRONICS CO., LTD.)

#### Cable side:

R04-P12F (manufactured by TAJIMI ELECTRONICS CO., LTD.) ...... Waterproof type R03-PB12F (manufactured by TAJIMI ELECTRONICS CO., LTD.) ..... Non-waterproof type

Pin arrangement	Input/output specifications			
	A/B signal output	Analog output		
A	A	+COS		
В	*A	-COS		
С	В	+SIN		
D	*B	-SIN		
E	Z	+REF		
F	*Z	-REF		
G	+5 V (power supply)			
Н	0 V (power supply)			
J	0 V (signal)			
K	0 V (signal)			
L	+5 VS			
M	0 VS			



- 0 V is the circuit ground, and it is not connected to the frame ground.
- $\bullet$  Make sure that the power supply voltage is 5 V DC  $\pm$  5 % at the input connector to the interface unit.
- +5 V S and 0 V S are for checking the voltage (remote sensing function) applied to the input connector of the interface unit.

These voltages can be used to check and control for drops in the supply voltage due to the cables.

When using a power supply that cannot control power supply fluctuations,

a power supply input terminal can be used to reduce the supply voltage drops occurring due to the cable length.

In this case, connect the cable to the respective +5 V or 0 V power supply.

- The appropriate cable thickness is AWG28 to AWG24.
- Connect all of the 0 V terminals to prevent mis-wiring.
- Use shielded cables for all cabling.
- Use twisted-pair cables for the output signals.

Use cables so that the following signals are paired: A and \*A, B and \*B, Z and \*Z, +SIN and -SiN, +COS and -COS, +REF and -REF.

#### Connectors (Open type) BL57-RE, BL57NE

	Input/output specifications		
Pin arrangement	A/B signal output (Output format F, G)	Analog output (Output format H)	
1	A	+COS	
2	*A	-COS	
3	В	+SIN	
4	*B	-SIN	
5	REF	(Not connectable)	
6	*REF	0 V (power supply)	
7	+5 V (power supply)	0VS	
8	ALM	(Not connectable)	
9	+5 V (power supply)	+5 V (power supply)	
10	*ALM	+5VS	
11	+5VS	+REF	
12	(Not connectable)	-REF	
13	+5 V (power supply)	(Not connectable)	
14	SIN (M)	(Not connectable)	
15	0 V (power supply)	(Not connectable)	
16	COS (M)		
17	0 V (power supply)		
18	(Not connectable)		
19	0VS		
20	(Not connectable)		
21	OV (M)		
22	(Not connectable)		
23	0 V (power supply)		
24	(Not connectable)		
25	0 V (signal)		
26	(Not connectable)		

#### Interface unit side:

A/B signal output : 10226-52A2JL (manufactured by SUMITOMO 3M Limited)

Analog output : D02-M15SAG-26L9

(manufactured by Japan Aviation Electronics Industry, Limited)

#### Cable side:

A/B signal output : Plug 10126-3000VE

(manufactured by SUMITOMO 3M Limited)

: Shell 10326-52F0-00S

(manufactured by SUMITOMO 3M Limited)

Analog output : Plug D02-M15PG-N-F0 (manufactured by Japan Aviation Electronics Industry, Limited)

: Contact When AWG24 wire is used

D02-22-22P-PKG100

(manufactured by Japan Aviation Electronics Industry, Limited)

: Contact When AWG26-28 wire is used

D02-22-26P-PKG100

(manufactured by Japan Aviation Electronics Industry, Limited)

(manufactured by Japan Aviation Electronics Industry, Limited)







http://www.mgscale.com

### Magnescale